

Quality and Compatibility Statement

Quality

MySling slings are manufactured to the highest standards and quality is closely monitored at all stages in the manufacturing process. Each sling is visually inspected before it is packed.

Before being released to the market, each design has been tested in accordance with the British Standards BS EN 10535:2006 'Hoists for the transfer of disabled persons' and BS EN 12182:2012 'Assistive products for persons with disability'.

Testing under these standards include:

- Ascertaining which manufacturer's hoists and spreader bars can be used in safe combination with the sling
- Weight testing (1.5 times the Safe Working Load must be statically held for 20 minutes)
- Fire retardancy testing (EN1021-1 and EN1021-2)
- Laundry tests to ascertain overall shrinkage

In addition, we carry out testing for our own internal quality procedures which far exceed the minimum requirements of the standard. This additional testing includes:

- Load testing when wet
- Overloading to breaking point

All products carry a CE (Conformity European) mark following guidelines set out by the Medicines and Healthcare products Regulatory Agency.

Compatibility

Certain hoist manufacturers have displayed a reluctance to recognise the use of alternative slings with their products. This may be motivated by commercial or competitive reasons, or by fear of litigation. In any event, a thorough written risk assessment conducted by a competent assessor is all that is required to ensure that an alternative sling provides a safe and appropriate solution for a client's specific need, most often at a significantly reduced cost. Further guidance on risk assessments and competency can be found under the Provision and Use of Work Equipment Regulations (PUWER 5 Dec 1998)

MySling slings are manufactured and are safe and fit for purpose, visibly marked with any relevant information required for a safe transfer, including information of the style of the replacement sling, the maximum weight limit, inspection instructions and unique serial numbers for identification.

We conduct regular research and testing to ensure that MySling replacement slings are safe to use in combination with hoists by the leading manufacturers in the UK market. Compatible lifts and standing aids

are listed in the table in Appendix I of this document. The suitability of any sling for both patient and carer is of course subject to an individual assessment carried out by a appropriately qualified individual.

Master Indemnity Agreements for trials and samples

Lisclare has entered into a Master Indemnity Agreement with the Department of Health which indemnifies NHS trusts, PCT's and their employees from liability to damage or injury caused by way of fault in the design or workmanship of our slings during trials or through the use of a sample sling.

Our Master Indemnity Agreement reference numbers are:

For use with form:	Lisclare Reference
A – Goods on loan from Lisclare to a trust	IFA822
B – For use with samples given to the trust by Lisclare	IFB822

Contact

The MySling range is designed, manufactured and distributed in the UK and Ireland by Lisclare.

Further information is available on our website at www.lisclare.com, or by contacting our sales support desk on 0151 495 5750.

Signed for and on behalf of Lisclare Limited:



Stephen Hall
Managing Director
Lisclare Limited

8 October 2020



Appendix I - Passive Lifts and Active Standing Aids compatible with the MySling range

Passive lifts with loop fixings

Manufacturer	Model
Arjo (with loop spreader bar)	Maxi/Maxi Sky range
Arjo (with loop spreader bar)	Opera
Arjo (with loop spreader bar)	Tenor
BHM	Overhead lifts
Chiltern	Wispa
Drive Medical (Days)	Casa range
Ergolet	Atlas
Ergolet	Luna
Ergolet	Selena
Ergolet	Solar
Etac (Molift)	Air
Etac (Molift)	Mover range
Etac (Molift)	Nomad
Etac (Molift)	Partner
Etac (Molift)	Smart
Guldmann	GH1
Guldmann	GH3
Guldmann	GL5
Handicare	Carina
Handicare	Eva
Handicare	Rise
Harvest Healthcare	Olympic
Hill-Rom (Liko)	Golvo
Hill-Rom (Liko)	Likorall
Hill-Rom (Liko)	M220
Hill-Rom (Liko)	M240
Hill-Rom (Liko)	Multirall
Hill-Rom (Liko)	Ultra Twin
Hill-Rom (Liko)	Uno
Hill-Rom (Liko)	Viking range
Humancare	Roomer
Invacare	Burdie range
Invacare	iTransia
Invacare	Robin
Joerns (Oxford)	Advance
Joerns (Oxford)	Calibre
Joerns (Oxford)	MaxiLift
Joerns (Oxford)	MidiLift

Manufacturer	Model
Joerns (Oxford)	MiniLift
Joerns (Oxford)	Stature
Joerns (Oxford)	Voyager
Joerns (Oxford)	Presence
Opemed	Opemed portable overheads
Opemed	OT200
Parkhouse	Parklite 180
Prism	A- range
Prism	E- range
Prism	Freeway
Prism	M- range
Prism	S- range
Prism	Transactive

Passive lifts with clip fixings

Manufacturer	Model
Arjo	Maxi range
Arjo	Maxi Sky range
Arjo	Bianca
Arjo	Trixie
Arjo	Dextra
Arjo	Marisa
Arjo	Opera
Arjo	Tempo

Active lifts with loop fixings (stand aids)

Manufacturer	Model
Hill-Rom (Liko)	Sabina
Invacare	Reliant 350
Handicare	Minilift range
Molift	Raiser
Joerns (Oxford)	Elevate
Prism	S range
Parkhouse	Parklite range

Active lifts with clip fixings (stand aids)

Manufacturer	Model
Arjo	Sara range
Arjo	Sarita